

# Gynetics Medical Products

## CERTIFICATE OF QUALITY

**ITEM name and code:**  
Follicle Aspiration Needle #4551 FAS

**LOT number:** 20180912

**STERILIZATION date:** 01-10-2018

**EXPIRY date:** 12-09-2022

**STERILIZATION technique:** Gamma Irradiation

**STERILIZATION dose:** 25 - 39 kGray

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



Greetje Hermans  
Gynetics Medical Products

**SIGNATURE date:**  
31-10-2018