

CERTIFICATE OF QUALITY

ITEM name and code #4219 Emtrac C-240 - Embryo Transfer Cathether
LOT NUMBER 20180923
EXPIRY DATE 23 September 2022
STERILIZATION DATE 7 October 2018
STERILIZATION TECHNIQUE Gamma Irradiation
STERILIZATION DOSE 22,5 - 45 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass

Greetje Hermans
Gynétics Medical Products N.V.

Signature date:
8 November 2018