

Gynetics Medical Products CERTIFICATE OF QUALITY

ITEM name and code: #4164 Probet
Endometrial Curette

LOT NUMBER: 20180926

EXPIRY DATE: 26 September 2022

STERILIZATION DATE: 01 October 2018

STERILIZATION TECHNIQUE: EO

**Results of sterility test with bio-
indicators** No growth of organisms

Gynetics guarantees the following results due to a validated EO sterilization process.

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATIONS	Result
Ethylene oxide residue testing	<250 ppm	Pass
Ethylene chlorhydrin residue testing	<250 ppm	Pass
Suction function testing	<-110 mbars	Pass



Greetje Hermans
Gynetics Medical Products N.V.

2-11-2018
Date