

CERTIFICATE OF QUALITY

ITEM name and code	#4219 Emtrac C-200 - Embryo Transfer Cathether
LOT NUMBER	20181024
EXPIRY DATE	24 October 2022
STERILIZATION DATE	26 October 2018
STERILIZATION TECHNIQUE	Gamma Irradiation
STERILIZATION DOSE	22,5 - 45 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Greetje Hermans
Gynetics Medical Products N.V.

Signature date:
21 November 2018