

## CERTIFICATE OF QUALITY

<b>ITEM name and code</b>	#4219 Emtrac A - Embryo Transfer Cathether
<b>LOT NUMBER</b>	20181107
<b>EXPIRY DATE</b>	7 November 2022
<b>STERILIZATION DATE</b>	11 November 2018
<b>STERILIZATION TECHNIQUE</b>	Gamma Irradiation
<b>STERILIZATION DOSE</b>	22,5 - 45 kGray

## QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Greetje Hermans  
Gynétics Medical Products N.V.

Signature date:  
29 November 2018