

## CERTIFICATE OF QUALITY

**ITEM name and code** #4551 FAS - Follicle Aspiration Needle  
**LOT NUMBER** 20181113  
**EXPIRY DATE** 13 November 2022  
**STERILIZATION DATE** 23 November 2018  
**STERILIZATION TECHNIQUE** Gamma Irradiation  
**STERILIZATION DOSE** 22,5 - 45 kGray

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass

Greetje Hermans  
Gynetics Medical Products N.V.

Signature date:  
12 December 2018