

## **CERTIFICATE OF QUALITY**

ITEM name and code #4164 Probet - Endometrial Suction Curette

**LOT NUMBER** 20181211

**EXPIRY DATE** 11 December 2022

STERILIZATION DATE 20 December 2018

STERILIZATION TECHNIQUE Ethylene Oxide

**RESULTS OF BI** No growth of organisms

## **QUALITY CONTROL TESTING PROCEDURES**

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass

Greetje Hermans

Signature date:

Gynétics Medical Products N.V.

9 January 2019