

CERTIFICATE OF QUALITY

| | |
|--------------------------------|--|
| ITEM name and code | #4000 The Tulip - Embryo Transfer Catheter |
| LOT NUMBER | 20181212 |
| EXPIRY DATE | 12 December 2022 |
| STERILIZATION DATE | 16 December 2018 |
| STERILIZATION TECHNIQUE | Gamma Irradiation |
| STERILIZATION DOSE | 22,5 - 45 kGray |

QUALITY CONTROL TESTING PROCEDURES

| TEST | SPECIFICATION | CONCLUSION |
|-----------------------|----------------------|-------------------|
| Endotoxin Assay (LAL) | <0.5 EU/device | Pass |

| TEST | SPECIFICATIONS | CONCLUSION |
|--------------------|-----------------------|-------------------|
| Mouse Embryo Assay | > 80% | Pass |
| MEA Control | > 80% | Pass |



Greetje Hermans
Gynetics Medical Products N.V.

Signature date:
10 January 2019