

CERTIFICATE OF QUALITY

ITEM name and code #4164 Probet - Endometrial Suction Curette

LOT NUMBER 20181212

EXPIRY DATE 12 December 2022

STERILIZATION DATE 20 December 2018

STERILIZATION TECHNIQUE Ethylene Oxide

RESULTS OF BI No growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass

Greetje Hermans

Signature date:

Gynétics Medical Products N.V.

9 January 2019