

## CERTIFICATE OF QUALITY

<b>ITEM name and code</b>	#4000 Tulip LS - Embryo Transfer Catheter
<b>LOT NUMBER</b>	20181219
<b>EXPIRY DATE</b>	19 December 2022
<b>STERILIZATION DATE</b>	23 December 2018
<b>STERILIZATION TECHNIQUE</b>	Gamma Irradiation
<b>STERILIZATION DOSE</b>	22,5 - 45 kGray

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Greetje Hermans  
Gynetics Medical Products N.V.

Signature date:  
24 January 2019