


CERTIFICATE OF QUALITY

ITEM name and code	#4219 Emtrac Delphin - Embryo Transfer Catheter
LOT NUMBER	20190116
EXPIRY DATE	16 January 2023
STERILIZATION DATE	19 January 2019
STERILIZATION TECHNIQUE	Ethylene Oxide
RESULTS OF BI	No growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Evelien Sommen
Gynétics Medical Products N.V.

Signature date:
28 February 2019