

CERTIFICATE OF QUALITY

ITEM name and code	#4000 The Tulip - Embryo Transfer Catheter
LOT NUMBER	20190116
EXPIRY DATE	16 January 2023
STERILIZATION DATE	19 January 2019
STERILIZATION TECHNIQUE	Gamma Irradiation
STERILIZATION DOSE	22,5 - 45 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Greetje Hermans
Gynetics Medical Products N.V.

Signature date:
13 February 2019