

CERTIFICATE OF QUALITY

ITEM name and code	#4225 Smooze Long OM - Intra-Uterine Insemination Catheter/Cannula Soft
LOT NUMBER	20190122
EXPIRY DATE	22 January 2023
STERILIZATION DATE	25 January 2019
STERILIZATION TECHNIQUE	Ethylene Oxide
RESULTS OF BI	No growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass

Evelien Sommen
Gynétics Medical Products N.V.

Signature date:
21 February 2019