


CERTIFICATE OF QUALITY

ITEM name and code	#2000 Semtrac W-180 SP - Soft Embryo Transfer Catheter
LOT NUMBER	20190213
EXPIRY DATE	13 February 2023
STERILIZATION DATE	17 February 2019
STERILIZATION TECHNIQUE	Ethylene Oxide
RESULTS OF BI	No growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Evelien Sommen
Gynétics Medical Products N.V.

Signature date:
13 March 2019