


CERTIFICATE OF QUALITY

ITEM name and code	#4219 Emtrac Delphin - Embryo Transfer Catheter
LOT NUMBER	20190306
EXPIRY DATE	6 March 2023
STERILIZATION DATE	11 March 2019
STERILIZATION TECHNIQUE	Ethylene Oxide
RESULTS OF BI	No growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Evelien Sommen
Gynétics Medical Products N.V.

Signature date:
3 April 2019