


CERTIFICATE OF QUALITY

ITEM name and code	#3100 Handling Pipette - Handling Pipette
LOT NUMBER	20190410
EXPIRY DATE	10 April 2023
STERILIZATION DATE	22 April 2019
STERILIZATION TECHNIQUE	Gamma Irradiation
STERILIZATION DOSE	22,5 - 45 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Evelien Sommen
Gynétics Medical Products N.V.

Signature date:
15 May 2019