

CERTIFICATE OF QUALITY

ITEM name and code	#4219 Emtrac Set - Embryo Transfer Catheter
LOT NUMBER	20190424
EXPIRY DATE	24 April 2023
STERILIZATION DATE	28 April 2019
STERILIZATION TECHNIQUE	Gamma Irradiation
STERILIZATION DOSE	22,5 - 45 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass
Evelien Sommen	Signature date:	
Gynétics Medical Products N.V.	22 May 2019	