

CERTIFICATE OF QUALITY

ITEM name and code	#4220 IUI Memo - Intra-Uterine Insemination Catheter/Cannula with memory
LOT NUMBER	20190508
EXPIRY DATE	8 May 2023
STERILIZATION DATE	12 May 2019
STERILIZATION TECHNIQUE	Gamma Irradiation
STERILIZATION DOSE	22,5 - 45 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Evelien Sommen
Gynetics Medical Products N.V.

Signature date:
29 May 2019