

CERTIFICATE OF QUALITY

ITEM name and code #4551 FAS - Follicle Aspiration Needle
LOT NUMBER 20190527
EXPIRY DATE 27 May 2023
STERILIZATION DATE 1 June 2019
STERILIZATION TECHNIQUE Gamma Irradiation
STERILIZATION DOSE 22,5 - 45 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Greetje Hermans
Gynetics Medical Products N.V.

Signature date:
19 June 2019