

## CERTIFICATE OF QUALITY

|                                |  |
|--------------------------------|--|
| <b>ITEM name and code</b>      | #4220 IUI Memo - Intra-Uterine Insemination Catheter/Cannula with memory |
| <b>LOT NUMBER</b>              | 20190619   |
| <b>EXPIRY DATE</b>             | 19 June 2023   |
| <b>STERILIZATION DATE</b>      | 23 June 2019   |
| <b>STERILIZATION TECHNIQUE</b> | Gamma Irradiation  |
| <b>STERILIZATION DOSE</b>      | 22,5 - 45 kGray  |

### QUALITY CONTROL TESTING PROCEDURES

| TEST                  | SPECIFICATION  | CONCLUSION |
|-----------------------|----------------|------------|
| Endotoxin Assay (LAL) | <0.5 EU/device | Pass       |
| TEST                  | SPECIFICATIONS | CONCLUSION |
| Mouse Embryo Assay    | > 80%          | Pass       |
| MEA Control           | > 80%          | Pass       |



Greetje Hermans

Gynétics Medical Products N.V.

Signature date:

10 July 2019