

CERTIFICATE OF QUALITY

ITEM name and code	#4219 Emtrac C-200 - Embryo Transfer Cathether
LOT NUMBER	20190619
EXPIRY DATE	19 June 2023
STERILIZATION DATE	23 June 2019
STERILIZATION TECHNIQUE	Gamma Irradiation
STERILIZATION DOSE	22,5 - 45 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass

Greetje Hermans
Gynétics Medical Products N.V.

Signature date:
10 July 2019