

CERTIFICATE OF QUALITY

ITEM name and code	#3125 Handling Pipette - Handling Pipette
LOT NUMBER	20190626
EXPIRY DATE	26 June 2023
STERILIZATION DATE	3 July 2019
STERILIZATION TECHNIQUE	Gamma Irradiation
STERILIZATION DOSE	22,5 - 45 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass

Greetje Hermans
Gynétics Medical Products N.V.

Signature date:
5 August 2019