

## CERTIFICATE OF QUALITY

<b>ITEM name and code</b>	#4219 Emtrac Delphin - Embryo Transfer Catheter
<b>LOT NUMBER</b>	20190626
<b>EXPIRY DATE</b>	26 June 2023
<b>STERILIZATION DATE</b>	29 June 2019
<b>STERILIZATION TECHNIQUE</b>	Ethylene Oxide
<b>RESULTS OF BI</b>	No growth of organisms

### QUALITY CONTROL TESTING PROCEDURES

<b>TEST</b>	<b>SPECIFICATION</b>	<b>CONCLUSION</b>
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
<b>TEST</b>	<b>SPECIFICATIONS</b>	<b>CONCLUSION</b>
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Greetje Hermans

Gynétics Medical Products N.V.

Signature date:

5 August 2019