

CERTIFICATE OF QUALITY

| | |
|--------------------------------|--|
| ITEM name and code | #4551 FAS - Follicle Aspiration Needle |
| LOT NUMBER | 20190910 |
| EXPIRY DATE | 10 September 2023 |
| STERILIZATION DATE | 25 September 2019 |
| STERILIZATION TECHNIQUE | Gamma Irradiation |
| STERILIZATION DOSE | 22,5 - 45 kGy |

QUALITY CONTROL TESTING PROCEDURES

| TEST | SPECIFICATION | CONCLUSION |
|---|----------------|------------|
| Endotoxin Assay (LAL) | <1.0 EU/device | Pass * |
| TEST | SPECIFICATIONS | CONCLUSION |
| Mouse Embryo Assay (MEA) <i>expressed as day 5 expanded blastocyst stage</i> | > 80% | Pass * |

* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan

Greetje Hermans

Gynétics Medical Products N.V.

Signature date:

30 October 2019