

## CERTIFICATE OF QUALITY

**ITEM name and code** #4164 Probet - Endometrial Suction Curette

**LOT NUMBER** 20190918

**EXPIRY DATE** 18 September 2023

**STERILIZATION DATE** 22 September 2019

**STERILIZATION TECHNIQUE** Ethylene Oxide

Biological Indicator testing: no growth of organisms

## QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *

\* LAL results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan

Greetje Hermans

Gynétics Medical Products N.V.

Signature date:

23 October 2019