

CERTIFICATE OF QUALITY

ITEM name and code	#4225 Smooze - Intra-Uterine Insemination Catheter/Cannula Soft
LOT NUMBER	20191113
EXPIRY DATE	13 November 2023
STERILIZATION DATE	17 November 2019
STERILIZATION TECHNIQUE	Ethylene Oxide
	Biological Indicator testing: no growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay (MEA) <i>expressed as day 5 expanded blastocyst stage</i>	> 80%	Pass *

* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan

Greetje Hermans

Gynétics Medical Products N.V.

Signature date:

2 January 2020