

CERTIFICATE OF QUALITY

ITEM name and code #4164 Probet - Endometrial Suction Curette

LOT NUMBER 20200108

EXPIRY DATE 8 January 2024

STERILIZATION DATE 17 January 2020

STERILIZATION TECHNIQUE Ethylene Oxide

Biological Indicator testing: no growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *

* LAL results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan

Greetje Hermans

Gynétics Medical Products N.V.

Signature date:

12 February 2020