

CERTIFICATE OF QUALITY

ITEM name and code	#4225 Smooze Long - Intra-Uterine Insemination Catheter/Cannula Soft
LOT NUMBER	20200219
EXPIRY DATE	19 February 2024
STERILIZATION DATE	21 February 2020
STERILIZATION TECHNIQUE	Ethylene Oxide
	Biological Indicator testing: no growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *

* LAL results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Greetje Hermans

Gynétics Medical Products N.V.

Signature date:

18 March 2020