

## CERTIFICATE OF QUALITY

<b>ITEM name and code</b>	#3100 Handling Pipette - Handling Pipette
<b>LOT NUMBER</b>	20200624
<b>EXPIRY DATE</b>	24 June 2024
<b>STERILIZATION DATE</b>	26 June 2020
<b>STERILIZATION TECHNIQUE</b>	Gamma Irradiation
<b>STERILIZATION DOSE</b>	22,5 - 45 kGy

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay (MEA) <i>expressed as day 5 expanded blastocyst stage</i>	> 80%	Pass *

\* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Merve Toprak  
Gynétics Medical Products N.V.

Signature date:  
17 July 2020