

CERTIFICATE OF QUALITY

ITEM name and code #4164 Probet - Endometrial Suction Curette
LOT NUMBER 20200818
EXPIRY DATE 18 August 2024
STERILIZATION DATE 23 August 2020
STERILIZATION TECHNIQUE Ethylene Oxide

Biological Indicator testing: no growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *

* LAL results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Merve Toprak
Gynétics Medical Products N.V.

Signature date:
18 September 2020