

CERTIFICATE OF QUALITY

ITEM name and code	#4219 Emtrac Set - Embryo Transfer Catheter
LOT NUMBER	20201118
EXPIRY DATE	18 November 2024
STERILIZATION DATE	21 November 2020
STERILIZATION TECHNIQUE	Gamma Irradiation
STERILIZATION DOSE	22,5 - 45 kGy

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay (MEA) <i>expressed as day 5 expanded blastocyst stage</i>	> 80%	Pass *

* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Merve Toprak
Gynetics Medical Products N.V.

Signature date:
18 December 2020