

CERTIFICATE OF QUALITY

ITEM name and code #2000 Semtrac W-180 - Soft Embryo Transfer Catheter
LOT NUMBER 20210310
EXPIRY DATE 10 March 2025
STERILIZATION DATE 23 April 2021
STERILIZATION TECHNIQUE Ethylene Oxide
 Biological Indicator testing: no growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay (MEA) <i>expressed as day 5 expanded blastocyst stage</i>	> 80%	Pass *

* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Merve Toprak
 Gynétics Medical Products N.V.

Signature date:
 25 May 2021