

## CERTIFICATE OF QUALITY

**ITEM name and code** #4164 Probet - Endometrial Suction Curette

**LOT NUMBER** 20210519

**EXPIRY DATE** 19 May 2025

**STERILIZATION DATE** 4 June 2021

**STERILIZATION TECHNIQUE** Ethylene Oxide

Biological Indicator testing: no growth of organisms

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *

\* LAL results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Merve Toprak  
Gynétics Medical Products N.V.

Signature date:  
6 July 2021