

CERTIFICATE OF QUALITY

ITEM name and code #4164 Probet - Endometrial Suction Curette

LOT NUMBER 20210623

EXPIRY DATE 23 June 2025

STERILIZATION DATE 6 July 2021

STERILIZATION TECHNIQUE Ethylene Oxide

Biological Indicator testing: no growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *

* LAL results are conform the internal quality control procedure in which testing is conducted based on a repres sampling plan



Merve Toprak
Gynétics Medical Products N.V.

Signature date:
9 August 2021