

## CERTIFICATE OF QUALITY

**ITEM name and code** #4219 Emtrac Delphin - Embryo Transfer Catheter

**LOT NUMBER** 20210825

**EXPIRY DATE** 25 August 2025

**STERILIZATION DATE** 31 August 2021

**STERILIZATION TECHNIQUE** Ethylene Oxide

Biological Indicator testing: no growth of organisms

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay (MEA) <i>expressed as day 5 expanded blastocyst stage</i>	> 80%	Pass *

\* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Merve Toprak  
Gynetics Medical Products N.V.

Signature date:  
28 September 2021