

## CERTIFICATE OF QUALITY

<b>ITEM name and code</b>	#4164 Probet - Endometrial Suction Curette
<b>LOT NUMBER</b>	20210914
<b>EXPIRY DATE</b>	14 September 2025
<b>STERILIZATION DATE</b>	22 September 2021
<b>STERILIZATION TECHNIQUE</b>	Ethylene Oxide
	Biological Indicator testing: no growth of organisms

## QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *

\* LAL results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Merve Toprak  
 Gynetics Medical Products N.V.

Signature date:  
 18 October 2021