

## CERTIFICATE OF QUALITY

|                                |  |
|--------------------------------|--|
| <b>ITEM name and code</b>      | #4551 FAS - Follicle Aspiration Needle |
| <b>LOT NUMBER</b>              | 20211026                               |
| <b>EXPIRY DATE</b>             | 26 October 2025                        |
| <b>STERILIZATION DATE</b>      | 3 November 2021                        |
| <b>STERILIZATION TECHNIQUE</b> | Gamma Irradiation                      |
| <b>STERILIZATION DOSE</b>      | 22,5 - 45 kGy                          |

### QUALITY CONTROL TESTING PROCEDURES

| TEST  | SPECIFICATION  | CONCLUSION |
|---|----------------|------------|
| Endotoxin Assay (LAL)   | <1.0 EU/device | Pass *     |
| TEST  | SPECIFICATIONS | CONCLUSION |
| Mouse Embryo Assay (MEA)<br><i>expressed as day 5 expanded blastocyst stage</i> | > 80%          | Pass *     |

\* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Merve Toprak  
Gynétics Medical Products N.V.

Signature date:  
27 January 2022