GYNÉTICS®
Practise better Medicine

CERTIFICATE OF QUALITY

ITEM name and code	#4220 IUI Memo - Intra-Uterine Insemination Catheter/Cannula with memory
LOT NUMBER	20220126
EXPIRY DATE	26 January 2026
STERILIZATION DATE	11 July 2022
STERILIZATION TECHNIQUE	Gamma Irradiation
STERILIZATION DOSE	22,5 - 45 kGy

QUALITY CONTROL TESTING PROCEDURES

TEST	**SPECIFICATION**	**CONCLUSION**
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *
TEST	**SPECIFICATIONS**	**CONCLUSION**
Mouse Embryo Assay (MEA) *expressed as day 5 expanded blastocyst stage*	> 80%	Pass *

* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan

Merve Toprak
Gynétics Medical Products N.V.

Signature date:
23 August 2022