

## CERTIFICATE OF QUALITY

**ITEM name and code** #4219 Emtrac Delphin - Embryo Transfer Catheter

**LOT NUMBER** 20220216

**EXPIRY DATE** 16 February 2026

**STERILIZATION DATE** 22 April 2022

**STERILIZATION TECHNIQUE** Ethylene Oxide

Biological Indicator testing: no growth of organisms

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay (MEA) <i>expressed as day 5 expanded blastocyst stage</i>	> 80%	Pass *

\* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Merve Toprak

Gynetics Medical Products N.V.

Signature date:

23 May 2022