

## CERTIFICATE OF QUALITY

**ITEM name and code** #4225 Smooze Long - Intra-Uterine Insemination Catheter/Cannula Soft

**LOT NUMBER** 20230125

**EXPIRY DATE** 25 January 2027

**STERILIZATION DATE** 5 February 2023

**STERILIZATION TECHNIQUE** Ethylene Oxide

Biological Indicator testing: no growth of organisms

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay (MEA)	> 80%	Pass *
<i>expressed as day 5 expanded blastocyst stage</i>		

\* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Merve Toprak  
Gynétics Medical Products N.V.

Signature date:  
28 February 2023