

CERTIFICATE OF QUALITY

ITEM name and code #4164 Probet - Endometrial Suction Curette

LOT NUMBER 20230308

EXPIRY DATE 8 March 2027

STERILIZATION DATE 26 March 2023

STERILIZATION TECHNIQUE Ethylene Oxide

Biological Indicator testing: no growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *

* LAL results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Merve Toprak
Gynétics Medical Products N.V.

Signature date:
2 May 2023