

CERTIFICATE OF QUALITY

ITEM name and code	#4225 Smooze Long - Intra-Uterine Insemination Catheter/Cannula Soft
LOT NUMBER	20231018
EXPIRY DATE	18 October 2027
STERILIZATION DATE	20 October 2023
STERILIZATION TECHNIQUE	Ethylene Oxide
	Biological Indicator testing: no growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay (MEA)	> 80%	Pass *
<i>expressed as day 5 expanded blastocyst stage</i>		

* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Merve Toprak
Gynétics Medical Products N.V.

Signature date:
13 November 2023