

## CERTIFICATE OF QUALITY

|                         |  |
|-------------------------|--|
| ITEM name and code      | #4220 IUI Memo - Intra-Uterine Insemination Catheter/Cannula with memory |
| LOT NUMBER              | 20231206   |
| EXPIRY DATE             | 6 December 2027  |
| STERILIZATION DATE      | 23 December 2023   |
| STERILIZATION TECHNIQUE | Gamma Irradiation  |
| STERILIZATION DOSE      | 22,5 - 45 kGy  |

## QUALITY CONTROL TESTING PROCEDURES

| TEST  | SPECIFICATION  | CONCLUSION |
|---|----------------|------------|
| Endotoxin Assay (LAL)                               | <1.0 EU/device | Pass *     |
| TEST  | SPECIFICATIONS | CONCLUSION |
| Mouse Embryo Assay (MEA)                            | > 80%          | Pass *     |
| <i>expressed as day 5 expanded blastocyst stage</i> |                |            |

\* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Merve Toprak  
Gynétics Medical Products N.V.

Signature date:  
25 January 2024