

CERTIFICATE OF QUALITY

ITEM name and code	#4164 Probet - Endometrial Suction Curette
LOT NUMBER	20231213
EXPIRY DATE	13 December 2027
STERILIZATION DATE	20 January 2024
STERILIZATION TECHNIQUE	Ethylene Oxide
	Biological Indicator testing: no growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *

* LAL results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Merve Toprak
Gynétics Medical Products N.V.

Signature date:
9 February 2024