

CERTIFICATE OF QUALITY

ITEM name and code #4219 Emtrac Delphin - Embryo Transfer Catheter

LOT NUMBER 20240313

EXPIRY DATE 13 March 2028

STERILIZATION DATE 22 March 2024

STERILIZATION TECHNIQUE Ethylene Oxide

Biological Indicator testing: no growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay (MEA)	> 80%	Pass *
<i>expressed as day 5 expanded blastocyst stage</i>		

* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Merve Toprak
Gynétics Medical Products N.V.

Signature date:
18 April 2024