

## CERTIFICATE OF QUALITY

<b>ITEM name and code</b>	#4225 Smooze - Intra-Uterine Insemination Catheter/Cannula Soft
<b>LOT NUMBER</b>	20240924
<b>EXPIRY DATE</b>	24 September 2028
<b>STERILIZATION DATE</b>	1 October 2024
<b>STERILIZATION TECHNIQUE</b>	Gamma Irradiation
	Biological Indicator testing: no growth of organisms

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay (MEA) <i>expressed as day 5 expanded blastocyst stage</i>	> 80%	Pass *

\* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Begum Gungormus

Gynétics Medical Products N.V.

Signature date:

16 October 2024