

CERTIFICATE OF QUALITY

ITEM name and code	#4164 Probet - Endometrial Suction Curette
LOT NUMBER	20240925
EXPIRY DATE	25 September 2028
STERILIZATION DATE	1 October 2024
STERILIZATION TECHNIQUE	Gamma Irradiation
	Biological Indicator testing: no growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *

* LAL results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Begum Gungormus
Gynétics Medical Products N.V.

Signature date:
16 October 2024