

## CERTIFICATE OF QUALITY

**ITEM name and code** #4219 Emtrac Delphin - Embryo Transfer Catheter

**LOT NUMBER** 20241120

**EXPIRY DATE** 20 November 2028

**STERILIZATION DATE** 26 November 2024

**STERILIZATION TECHNIQUE** Ethylene Oxide

Biological Indicator testing: no growth of organisms

## QUALITY CONTROL TESTING PROCEDURES

| TEST                                                | SPECIFICATION  | CONCLUSION |
|-----------------------------------------------------|----------------|------------|
| Endotoxin Assay (LAL)                               | <1.0 EU/device | Pass *     |
| TEST                                                | SPECIFICATIONS | CONCLUSION |
| Mouse Embryo Assay (MEA)                            | > 80%          | Pass *     |
| <i>expressed as day 5 expanded blastocyst stage</i> |                |            |

\* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Eef Wouters  
Gynétics Medical Products N.V.

Signature date:  
17 December 2024