

## CERTIFICATE OF QUALITY

|                                |   |
|--------------------------------|---|
| <b>ITEM name and code</b>      | #3125 Handling Pipette - Handling Pipette |
| <b>LOT NUMBER</b>              | 20250121                                  |
| <b>EXPIRY DATE</b>             | 21 February 2029                          |
| <b>STERILIZATION DATE</b>      | 26 January 2025                           |
| <b>STERILIZATION TECHNIQUE</b> | Gamma Irradiation                         |
| <b>STERILIZATION DOSE</b>      | 22,5 - 45 kGy                             |

### QUALITY CONTROL TESTING PROCEDURES

| TEST  | SPECIFICATION  | CONCLUSION |
|---|----------------|------------|
| Endotoxin Assay (LAL)                               | <1.0 EU/device | Pass *     |
| TEST  | SPECIFICATIONS | CONCLUSION |
| Mouse Embryo Assay (MEA)                            | > 80%          | Pass *     |
| <i>expressed as day 5 expanded blastocyst stage</i> |                |            |

\* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Eef Wouters  
Gynétics Medical Products N.V.

Signature date:  
21 February 2025