

CERTIFICATE OF QUALITY

ITEM name and code #4164 Probet - Endometrial Suction Curette

LOT NUMBER 20250312

EXPIRY DATE 12 March 2029

STERILIZATION DATE 18 March 2025

STERILIZATION TECHNIQUE Ethylene Oxide

Biological Indicator testing: no growth of organisms

QUALITY CONTROL TESTING PROCEDURES

| TEST | SPECIFICATION | CONCLUSION |
|---|----------------|------------|
| Endotoxin Assay (LAL) | <1.0 EU/device | Pass * |
| TEST | SPECIFICATIONS | CONCLUSION |
| Mouse Embryo Assay (MEA) | > 80% | Pass * |
| <i>expressed as day 5 expanded blastocyst stage</i> | | |

* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Eef Wouters

Gynétics Medical Products N.V.

Signature date:

4 April 2025